

Calibration Certificate

Accredited Calibration



Cal Cert ID: 194433

Page 1 of 2

Cust. Name: WASHINGTON STATE PATROL

Address: 11018 NE 51ST CIRCLE

City: VANCOUVER

State/Zip: WA 98682 US

Instr SN: 410004SV/651311

Out1 Min Range: 8.0000 psi

Cal Date: 04/28/2021

Instr Descr: CPG 2400

Out1 Max Range: 17.0000 psi

Cal Time: 12:21 pm

Sensor SN: 410004SV/651311

Limit of Error: 0.030% R

Issue Date: 04/28/2021

Sensor Descr: Silicon Pressure Transducer

Cal Tech: JUDYD

Procedure ID: WI02063

Pressure Type: Absolute

Order Nbr: 38452

Test Points Values and Readings

Test Point #	Reference Value PSI	Measured Values psi	Error % R	Uncertainty psi
1	8.00189	8.0018	-0.0012	1.9E-4
2	9.80172	9.8019	0.0018	2.2E-4
3	11.60153	11.6015	-0.0003	2.6E-4
4	13.40133	13.4012	-0.0010	3.0E-4
5	15.20118	15.2011	-0.0005	3.4E-4
6	17.00109	17.0008	-0.0017	3.8E-4
7	15.20115	15.2012	0.0003	3.4E-4
8	13.40133	13.4013	-0.0002	3.0E-4
9	11.60145	11.6015	0.0004	2.6E-4
10	9.80163	9.8016	-0.0003	2.2E-4
11	8.00179	8.0019	0.0013	1.9E-4

Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD 8000	10017/10016/9994	DH CPD 8000	CAL ID: 159795	12/11/2021

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

Remarks and Notes

As Found Calibration.

As Left Calibration.

Recalibration Date: 04/28/2022


An asterisk (*) denotes a measured condition where the error exceeds the maximum permissible error.

Dash numbers preceded by the sensor SN reference a primary (-1) or turndown (-2) range as applicable.

The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 and Mensor Management Manual MM 3.0, Revision C, dated November 18, 2020.

Metrological traceability is established through accreditation by the American Association for Laboratory Accreditation (A2LA), a signatory subject to the International Laboratory Accreditation Cooperation (ILAC) Arrangement as described in Annex A of ISO 17025:2017. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.

The results reported herein relate only to the item calibrated. This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.





Cal Cert ID: 194433

Page 2 of 2

Transducer was calibrated in a horizontal position.
Prior to the as-found calibration, the zero was adjusted 0.0021 psi.

I, Judy Dosal, certify the accuracy of this Calibration Certificate:

A handwritten signature in cursive script, appearing to read "Judy Dosal", written in dark ink.

Calibration Lab Technician

Mensor Calibration Certificates and Stickers

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments.

The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product.

For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.