

# Calibration Certificate

Accredited Calibration



Cal Cert ID: 240897

Cust. Name: WASHINGTON STATE PATROL  
Address: 811 EAST ROANOKE  
City: SEATTLE  
State/Zip: WA 98102 USA

Page 1 of 2

Instr SN: 410004SZ/651315 Out1 Min Range: 8.0000 psi  
Instr Descr: CPG2400 Out1 Max Range: 17.0000 psi  
Sensor SN: 410004SZ/651315 Limit of Error: 0.030% R  
Sensor Descr: Silicon Pressure Transducer  
Procedure ID: WI02063 Pressure Type: Absolute

Cal Date: 01/26/2024  
Cal Time: 1:21 am  
Issue Date: 01/26/2024  
Cal Tech: JODYP  
Order Nbr: 53325

## Test Points Values and Readings

Test Point #	Reference Value psi	Measured Values psi	Error % R	Uncertainty psi
1	8.00056	8.0010	0.0055	1.9E-4
2	9.80039	9.8011	0.0071	2.2E-4
3	11.60027	11.6010	0.0066	2.6E-4
4	13.40013	13.4006	0.0035	3.0E-4
5	15.19997	15.2012	0.0081	3.4E-4
6	16.99994	17.0009	0.0055	3.8E-4
7	15.19998	15.2011	0.0076	3.4E-4
8	13.40009	13.4009	0.0060	3.0E-4
9	11.60020	11.6013	0.0093	2.6E-4
10	9.80036	9.8007	0.0034	2.2E-4
11	8.00052	8.0010	0.0054	1.9E-4

## Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD8000	10017/10016/9994	DH CPD 8000	CAL ID: 205428	12/14/2024

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

## Remarks and Notes

As Found Calibration.

As Left Calibration.

Recalibration Date: 01/26/2025

The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 and Mensor Management Manual MM 3.0, Revision D, dated September 21, 2021.

Metrological traceability is established through accreditation by the American Association for Laboratory Accreditation (A2LA), a signatory subject to the International Laboratory Accreditation Cooperation (ILAC) Arrangement as described in Annex A of ISO 17025:2017. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.

The results reported herein relate only to the item calibrated. This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.

Transducer was calibrated in a horizontal position.

Prior to the as-found calibration, the zero was adjusted 0.001822 psi.

*Handwritten signature and date:*  
2/5/2023



Cal Cert ID: 240897

Page 2 of 2

I, Jody Pena, certify the accuracy of this Calibration Certificate:

  
\_\_\_\_\_  
Calibration Lab Technician

#### **Mensor Calibration Certificates and Stickers**

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments.

The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product.

For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.