

Calibration Certificate

Accredited Calibration



Cal Cert ID: 195591

Cust. Name: WASHINGTON STATE PATROL
Address: 811 EAST ROANOKE
City: SEATTLE
State/Zip: WA 98102 US

Page 1 of 2

Instr SN: 4100060E/651365 Out1 Min Range: 8.0000 psi
Instr Descr: CPG 2400 Out1 Max Range: 17.0000 psi
Sensor SN: 4100060E/651365 Limit of Error: 0.030% R
Sensor Descr: Silicon Pressure Transducer
Procedure ID: WI02063 Pressure Type: Absolute

Cal Date: 05/25/2021
Cal Time: 2:22 pm
Issue Date: 05/25/2021
Cal Tech: SALLYS
Order Nbr: 38943

Test Points Values and Readings

Test Point #	Reference Value PSI	Measured Values psi	Error % R	Uncertainty psi
1	8.00191	8.0019	-0.0001	1.9E-4
2	9.80178	9.8019	0.0013	2.2E-4
3	11.60155	11.6018	0.0021	2.6E-4
4	13.40146	13.4017	0.0018	3.0E-4
5	15.20127	15.2016	0.0022	3.4E-4
6	17.00122	17.0016	0.0022	3.8E-4
7	15.20124	15.2016	0.0023	3.4E-4
8	13.40140	13.4018	0.0030	3.0E-4
9	11.60148	11.6018	0.0028	2.6E-4
10	9.80169	9.8020	0.0031	2.2E-4
11	8.00188	8.0018	-0.0010	1.9E-4

Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD 8000	10017/10016/9994	DH CPD 8000	CAL ID: 159795	12/11/2021

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

Remarks and Notes

Recalibration Date: 05/25/2022.

As Found Calibration.

As Left Calibration.

The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 and Mensor Management Manual MM 3.0, Revision C, dated November 18, 2020.

Metrological traceability is established through accreditation by the American Association for Laboratory Accreditation (A2LA), a signatory subject to the International Laboratory Accreditation Cooperation (ILAC) Arrangement as described in Annex A of ISO 17025:2017. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.

The results reported herein relate only to the item calibrated. This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.

Transducer was calibrated in a horizontal position.

Prior to the as-found calibration, the zero was adjusted 0.0015 psi.



Cal Cert ID: 195591

Page 2 of 2

I, Sally Salazar, certify the accuracy of this Calibration Certificate:

Sally Salazar

Calibration Lab Technician

Mensor Calibration Certificates and Stickers

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments.

The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product.

For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.