

Calibration Certificate

Accredited Calibration



Cal Cert ID: 212618

Cust. Name: WASHINGTON STATE PATROL
Address: 2715 RUDKIN ROAD
City: UNION GAP
State/Zip: WA 98903 US

Page 1 of 2

Instr SN: 4100060F/651366 Out1 Min Range: 8.0000 psi
Instr Descr: CPG2400 Out1 Max Range: 17.0000 psi
Sensor SN: 4100060F/651366 Limit of Error: 0.030% R
Sensor Descr: Silicon Pressure Transducer
Procedure ID: WI02063 Pressure Type: Absolute

Cal Date: 06/17/2022
Cal Time: 9:49 am
Issue Date: 06/17/2022
Cal Tech: PHILIPR
Order Nbr: 44682

Test Points Values and Readings

Test Point #	Reference Value PSI	Measured Values psi	Error % R	Uncertainty psi
1	8.00026	8.0001	-0.0020	1.9E-4
2	9.80010	9.8004	0.0031	2.2E-4
3	11.60004	11.6003	0.0022	2.6E-4
4	13.39984	13.4004	0.0042	3.0E-4
5	15.19967	15.2004	0.0048	3.4E-4
6	16.99973	17.0003	0.0034	3.8E-4
7	15.19979	15.2004	0.0040	3.4E-4
8	13.39981	13.4005	0.0052	3.0E-4
9	11.59994	11.6003	0.0031	2.6E-4
10	9.80013	9.8003	0.0017	2.2E-4
11	8.00026	8.0003	0.0005	1.9E-4

Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD8000	10017/10016/9994	DH CPD 8000	CAL ID: 205428	12/14/2024

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

Remarks and Notes

As Found Calibration.

As Left Calibration.

RECALIBRATION DATE: 06/17/2023

An asterisk denotes a measurement error on a calibration that exceeds the limit of error (Tolerance Limit) based on simple acceptance regardless of uncertainty.

Dash numbers preceded by the sensor SN reference a primary (-1) or turndown (-2) range.

The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 and Mensor Management Manual MM 3.0, Revision D, dated September 21, 2021.

Metrological traceability is established through accreditation by the American Association for Laboratory Accreditation (A2LA), a signatory subject to the International Laboratory Accreditation Cooperation (ILAC) Arrangement as described in Annex A of ISO 17025:2017. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.

07-18-22

Cal Cert ID: 212618

Page 2 of 2

The results reported herein relate only to the item calibrated. This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.

Transducer was calibrated in a horizontal position.

Prior to the as-found calibration, the zero was adjusted 0.0032 psi.

I, Philip Romero, certify the accuracy of this Calibration Certificate:

A handwritten signature in dark ink, appearing to read "Philip Romero", is written over a horizontal line.

Calibration Lab Technician

Mensor Calibration Certificates and Stickers

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments.

The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product.

For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.