

# Calibration Certificate

Accredited Calibration



Cal Cert ID: 205212

Page 1 of 2

Cust. Name: WASHINGTON STATE PATROL - 6403

Address: 6403 W. ROWAND ROAD

City: SPOKANE

State/Zip: WA 99224 US

Instr SN: 4100060J/651370

Out1 Min Range: 8.0000 psi

Cal Date: 12/08/2021

Instr Descr: CPG2400

Out1 Max Range: 17.0000 psi

Cal Time: 8:01 am

Sensor SN: 4100060J/651370

Limit of Error: 0.030% R

Issue Date: 12/08/2021

Sensor Descr: Silicon Pressure Transducer

Cal Tech: SALLYS

Procedure ID: WI02063

Pressure Type: Absolute

Order Nbr: 41905

## Test Points Values and Readings

Test Point #	Reference Value PSI	Measured Values psi	Error % R	Uncertainty psi
1	8.00285	8.0028	-0.0006	1.9E-4
2	9.80268	9.8032	0.0053	2.2E-4
3	11.60247	11.6033	0.0071	2.6E-4
4	13.40237	13.4035	0.0084	3.0E-4
5	15.20224	15.2035	0.0083	3.4E-4
6	17.00218	17.0035	0.0078	3.8E-4
7	15.20228	15.2036	0.0087	3.4E-4
8	13.40234	13.4034	0.0079	3.0E-4
9	11.60250	11.6032	0.0060	2.6E-4
10	9.80262	9.8031	0.0049	2.2E-4
11	8.00281	8.0029	0.0011	1.9E-4

## Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD8000	10017/10016/9994	DH CPD 8000	CAL ID: 159795	12/11/2021

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

## Remarks and Notes

Recalibration Date: 12/08/2022.

As Found Calibration.

As Left Calibration.

The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 and Mensor Management Manual MM 3.0, Revision C, dated November 18, 2020.

Metrological traceability is established through accreditation by the American Association for Laboratory Accreditation (A2LA), a signatory subject to the International Laboratory Accreditation Cooperation (ILAC) Arrangement as described in Annex A of ISO 17025:2017. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.

The results reported herein relate only to the item calibrated. This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.

Transducer was calibrated in a horizontal position.

Prior to the as-found calibration, the zero was adjusted 0.0025 psi.

12/16/21 *[Signature]*



Cal Cert ID: 205212

Page 2 of 2

I, Sally Salazar, certify the accuracy of this Calibration Certificate:

*Sally Salazar*

Calibration Lab Technician

#### **Mensor Calibration Certificates and Stickers**

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments. The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product. For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.