WASHINGTON STATE PATROL

Report of Calibration

Guth-Eutechnics Digital Thermometer

Guth Part Number: GU-4300-DT

Serial Number: 091796

The Digital Thermometer calibration was verified at 30.00°C, 34.00°C, and 37.00°C by comparison to temperature standards traceable to the National Institute of Standards & Technology.

Simulator Temperature Degrees C	Master Digital Thermometer S/N 082708	Digital Thermometer Reading	Test Date
30.00	30.00°C	30.00°C	04/25/02
34.00	34.00°C	34.00°C	04/25/02
37.00	37.00°C	37.00°C	04/25/02

Tested by: _____

Date: _April 25, 2002_

Bostec Inc.