



Calibration Certificate

Accredited Calibration



Cal Cert ID: 145291

Page 1 of 2

Cust. Name: WASHINGTON STATE PATROL - 6403
Address: 6403 W. ROWAND ROAD
City: SPOKANE
State/Zip: WA 99224 US

Instr SN: 4100060J/651370 Out1 Min Range: 8.0000 PSI Cal Date: 01/03/2018
Instr Descr: CPG 2400 Out1 Max Range: 17.0000 PSI Cal Time: 10:24 pm
Sensor SN: 4100060J/651370 Limit of Error: 0.000 %F.S.+ 0.030 % R Cal Tech: SAMD
Procedure ID: WI02063 Pressure Type: Absolute Order Nbr: 20920

Test Points Values and Readings

Test Point #	Reference Value PSI	Measured Values PSI	Error % R	Uncertainty PSI
1	8.00166	8.0016	-0.001	.00018
2	9.80146	9.8016	0.001	.00022
3	11.60127	11.6018	0.005	.00026
4	13.40108	13.4016	0.004	.00029
5	15.20096	15.2015	0.004	.00033
6	17.00085	17.0014	0.003	.00037
7	15.20090	15.2014	0.003	.00033
8	13.40107	13.4016	0.004	.00029
9	11.60126	11.6017	0.004	.00026
10	9.80143	9.8015	0.001	.00022
11	8.00157	8.0014	-0.002	.00018

Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD 8000	10017/10016/9994	DH CPD 8000	CAL ID: 119787	12/10/2018

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

Remarks and Notes

As Found Calibration.
As Left Calibration.
Prior to "AS FOUND" calibration, zero was adjusted +.0016 PSI
Recalibration Date: 01/03/2019
An asterisk (*) denotes a measured condition where the error exceeds the maximum permissible error.
Dash numbers preceded by the sensor SN reference a primary (-1) or turndown (-2) range as applicable.
The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2005, ANSI/NCSL Z540-1 and Mensor Quality Manual 2.0, Revision R, dated March 10, 2015.
Standards used in this calibration are traceable to SI units of measurements through N.I.S.T. or recognized national metrology institutes. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.
This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.

Cal Cert ID: 145291

Page 2 of 2

I, Sam DelaRosa, certify the accuracy of this Calibration Certificate:

A handwritten signature in black ink, appearing to read "Sam DelaRosa", is positioned above a horizontal line.

Calibration Lab Technician

Mensor Calibration Certificates and Stickers

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments.

The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product.

For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.