



## Certificate of Calibration Everett Service Center

ISO 9001:2008 (10100/2)

Certificate Number: 649084-70530850:1382193983

<b>Result Summary:</b> PASS	<b>Date of Calibration:</b> 19 October 2013
<b>Data Type:</b> FOUND-LEFT	<b>Recommended Due Date:</b> 19 October 2014
<b>Manufacturer:</b> FLUKE	<b>Date of Certificate:</b> 19 October 2013
<b>Model:</b> 70 III	<b>Received Date:</b> 15 October 2013
<b>Serial Number:</b> 70530850	<b>Temperature:</b> 23.3 °C
<b>Description:</b> MULTIMETER	<b>Relative Humidity:</b> 38 %RH
<b>Procedure Name:</b> FLUKE 70-3: (1 YEAR) CAL VER/ALT 5520A	<b>Procedure Revision:</b> 1.0

**Customer Name:** STATE OF WASHINGTON  
**City, State:** SEATTLE, WA  
**Customer Item ID:**  
**PO Number:** WSP CCS 14-OCT-2013  
**RMA Number:** 30390931

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes, radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Rev 116 08/12 and Fluke Customer Support Services QAM 400 Rev. 002 03/22/2012.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

A handwritten signature in black ink, appearing to read 'Corey Dinardi', written over a horizontal line.

Corey Dinardi  
Metrology Technician

**FLUKE** ®

Certificate Number:  
649084-70530850:1382193983

Calibration Date:  
19-Oct-13

**Standards Used**

Asset #	Instrument Model	Cal Date	Cal Due
15267	FLUKE 5520A/1GHZ 120 CALIBRATOR	20 August 2013	20 May 2014

End of Report