



CERTIFICATE OF ANALYSIS

This is to certify that the product listed below was manufactured and inspected in accordance with the US Food and Drug Administration (FDA), ISO 13485 Quality System Requirements, and Covidien specifications; the subject product complies with all requirements. Further, the product is listed with the FDA and is manufactured at a facility which is registered with the FDA. All original records are maintained at the manufacturing facility.

Product: TT – BCS Gra 16 x 100 10mL P.O. - F

Item Code: 8881352796

Lot Number: 206853

Manufacture Date: April 20, 23-25, 2012

Expiration Date: 2014-03

Country of Origin: United States of America

Tested For	Acceptance Range	Results	Reference
Sodium Fluoride	Min .0880 g/tube Max .1120 g/tube	Min .0959 g/tube Max .1117 g/tube	FS10026251
Potassium Oxalate	Min .0150 g/Tube Max.0260 g/Tube	Min.0188 g/Tube Max.0223 g/Tube	FS10026251
H2O Draw	Min 69 mm/ Max 73 mm	Min 69 mm/Max 73 mm	FS10015185
Stopper Removal	Min 4.0 lbs.	Min 5.30 lbs.	GS10001670
Assay Potassium Oxalate	Spec.15.0-26.0mg/tube	Min. 19.40 mg/tube Max. 24.40 mg/tube	FS10026251
Assay Sodium Fluoride	Spec.88.0-112.0mg/tube	Min. 96.7 mg/tube Max. 109.0 mg/tube	FS10026251

Sterilized by Gamma Radiation - This certifies that sterilization records for the product listed on this certificate have been reviewed and found to be in compliance with the current Covidien requirements for sterility. These products have been processed in a validated sterilization cycle which meets the current standards and recommended practices in accordance with ANSI/AAMI/ISO standards.

Robin Rodelle
Quality Assurance

Date

COMMONWEALTH OF MASSACHUSETTS
Bristol County

On this 11th day of January, 2013, before me, the undersigned notary public, personally appeared Esie Cabral, proved to me through satisfactory evidence of identification, which were personal knowledge, to be the person whose name is signed on the preceding or attached document, and acknowledged to me that she signed it voluntarily for its stated purpose.

Notary Public

My commission expires: NOV 3, 2017