



December 5, 2024

Washington State Patrol
Toxicology Laboratory Division
2203 Airport Way S
Seattle WA, 98134

RE: Washington State Blood Alcohol Collection Kits

Blood Alcohol Collection Kits
Reorder Number: BA-2WA

TTF Lot No.	Kit Serial No.	Kit Exp.	Tube Lot No.	Tube Exp.	Prep Pad Exp.
93307	N/A	7/31/26	427160	7/31/26	N/A

Cardinal Health Medical Products

Certificate of Analysis



Cardinal Health products are designed, manufactured and inspected in accordance with product specific specifications in a controlled environment in accordance with established manufacturing guidelines and/or the applicable US Code of Federal Regulations 21CFR Part 820 "Quality System Regulation"; International Standard EN ISO 13485 "Medical Devices – Quality Management Systems" when applicable. Further, the products are manufactured in facilities that are appropriately registered with and have undergone applicable Agency and Notified Body audits against these requirements. All required manufacturing records are maintained in the Device History Record and are stored within a controlled documentation system.

Cardinal Health Product #	8881352796
Product Description	TT-BCS GRA 16X100 10 P.O.F.

Lot Number	Manufacture Date MM-DD-YYYY	Expiration Date MM-DD-YYYY	Test Type	Acceptance Criteria	Results
427160	8/20/2024	7/31/2026	Functional Testing	Product must meet applicable functional testing requirements	Meets specifications
			Visual Inspection	Product must meet applicable visual testing requirements, including labeling review	Meets specifications

The listed products were manufactured, inspected and released in accordance with established specifications for quality and efficacy. Product meets all functional and visual specifications.

Sign: 

Date: 12/5/2024

Print name: Keegan Chamberlain

Title: Director, Global Logistics Quality Management

If you have need for additional clarification or certificates, please navigate to the Cardinal Health certificates webtool at <https://certificates.cardinalhealth.com/> or contact your local Cardinal Health representative/customer service. For US customers only, contact customer service at 800.964.5227.

Cardinal Health Medical Products

Certificate of Conformance



Cardinal Health products are manufactured, inspected, tested and meet all requirements in accordance with our standards, Processing Factory Quality Procedures, Materials and Product specifications as well as in accordance with the ID FDA Quality System Regulations (21CFR Part 820) and ISO 13485 when applicable. Test reports are maintained on file and available for examination to demonstrate conformance to applicable specification requirements.

Cardinal Health Product #	8881352796
Product Description	TT-BCS GRA 16X100 10 P.O.F.

Lot Number	Manufacture Date MM-DD-YYYY	Expiration Date MM-DD-YYYY
427160	8/20/2024	7/31/2026

The listed products were manufactured, inspected and released in accordance with established specifications for quality and efficacy. Product meets all functional and visual specifications.

Sign: 

Date: 12/5/2024

Print name: Keegan Chamberlain

Title: Director, Global Logistics Quality Management

If you have need for additional clarification or certificates, please navigate to the Cardinal Health certificates webtool at <https://certificates.cardinalhealth.com/> or contact your local Cardinal Health representative/customer service. For US customers only, contact customer service at 800.964.5227.

Cardinal Health Medical Products

Certificate of Sterility



This certifies that products labeled as sterile have been processed in a validated sterilization cycle which meets the current standards and recommended published practices (ANSI/AAMI/ISO); sterilization records have been reviewed to ensure compliance with sterility specifications.

Cardinal Health Product #	8881352796
Product Description	TT-BCS GRA 16X100 10 P.O.F.

Lot Number	Manufacture Date MM-DD-YYYY	Expiration Date MM-DD-YYYY	Sterilization Method
427160	8/20/2024	7/31/2026	GAMMA

Sign: 

Date: 12/5/2024

Print name: Keegan Chamberlain

Title: Director, Global Logistics Quality Management

If you have need for additional clarification or certificates, please navigate to the Cardinal Health certificates webtool at <https://certificates.cardinalhealth.com/> or contact your local Cardinal Health representative/customer service. For US customers only, contact customer service at 800.964.5227.

© 2024 Cardinal Health. All Rights Reserved. CARDINAL HEALTH and the Cardinal Health LOGO are trademarks of Cardinal Health and may be registered in the US and/or in other countries. All other marks are the property of their respective owners.



TUG10 OEM Statement

The Tri-Tech Forensics, Inc. (TTF) gray stoppered 10ml blood tubes, product number TUG10 are made under TTF FDA 510(K) number K922342. These tubes are privately labeled and manufactured by Cardinal Health solely for Tri-Tech Forensics and are Cardinal Health product number 8881352796.

Tube Specifications

10ml – 16x100 mm

Potassium Oxalate = 15.0 - 26.0 mg/tube

Sodium Fluoride = 88.0 – 112.0 mg/tube

Sincerely

A handwritten signature in blue ink that reads "Eric A. Barton".

Eric Barton
Vice President

