

Kendall

www.tycohealthcare.com

RECEIVED

AUG 20 2009

WSP TOX LAB

To: Ms. Jennifer Rice
Washington State Toxicology Lab
2203 Airport Way S. Ste 360
Seattle, WA 98134

Fax: 206-262-6145

CERTIFICATE OF COMPLIANCE

This is to certify that the products listed below are in compliance with the current FDA Quality System Requirements (QSR) as stipulated in 21 CFR Part 820. Representative product was inspected and tested in accordance with current Kendall specifications and quality requirements.

<u>Product #</u>	<u>Description</u>	<u>Lot #</u>
8881352788	BCS GRA 16X100 10ML P.O+F	907910

This tube was designed for laboratory procedures requiring plasma or whole blood and chemistry procedures where glycoltic inhibition of the specimen is required. The tube was manufactured to the following specifications:

	Range	Nominal
Potassium Oxalate	15.0-26.0mg	20mg
Sodium Fluoride	20.0-32.0mg	25mg
Draw volume	9.0-11.0ml	10ml

David A. Olson

David A. Olson
Vice President Regulatory Affairs

On this 18th day of August, before me the undersigned notary public, personally appeared David A. Olson, proved to me through satisfactory evidence of identification, which were personal knowledge to be the person whose name is signed on the proceeding or attached document in my presence.

Maureen A. Burke
Notary Public
My commission expires: January 10, 2014

Date: August 13, 2009



MAUREEN A. BURKE
Notary Public
Commonwealth of Massachusetts
My Commission Expires
January 10, 2014