

COMPANY STATEMENT



July 29, 2021

BD Statement on False Recall Notice for BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations

Ensuring the safety and quality of our products is our top priority at BD. As part of that commitment, BD has a robust quality system in place, including a comprehensive post-market surveillance process. BD was recently notified by a customer that a recall announcement, purporting to be from BD, has been circulated. The falsified notice is dated June 4, 2021 and appears to be recalling one lot of BD Vacutainer® Fluoride Tubes, but was not issued by BD. **To be clear, there are no current recall notices for BD Vacutainer® Fluoride Tubes from BD.** BD did have a recall in 2019 for BD Vacutainer® Fluoride Tubes, but this recall was for a different, now expired, lot number and the recall has since been terminated by the FDA. The falsified recall notice includes many of the same elements from the 2019 recall notice but lists a different lot number and expiration date. We are investigating the situation and will take appropriate actions, including notifying regulatory bodies and law enforcement as needed.

For any further questions, please contact BD's customer response team at productcomplaints@bd.com or 1-844-8-BD-LIFE (844-823-5433).



STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2027 • (206) 262-6100 • FAX (206) 262-6145

July 27, 2021

Washington Association of Prosecuting Attorneys
206 10th Ave SE
Olympia, WA 98501

Subject: BD Vacutainer Recall Notification

On July 26, 2021, the Washington State Patrol Toxicology Laboratory received notification from manufacturer BD regarding a recall of BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations, catalog number 367001, lot number 8215683, distributed by BD beginning March 31, 2019. The *BD Urgent Medical Device Recall* notice, dated June 4, 2021, is attached.

The Toxicology Laboratory has contacted BD and Tritech Forensics, the distributor of the BD tubes, for additional information on this recall.

Sincerely,

Elizabeth Gough
Acting Toxicology Laboratory Division Commander
Washington State Patrol
Toxicology Laboratory Division

EG:eg/fc/ab

Please see July 29, 2021 BD Statement regarding Falsified June 4, 2021 Recall Notice



URGENT MEDICAL DEVICE RECALL

BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations

June 04, 2021

Product	Catalog Number	Lot Number	UDI (GTIN, DI + PI)	Exp. Date
BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations	367001	8215683	(01)30382903670018(17)200731 (10)8187663(30)0100	2017/9/30

For the Attention of: Lab Director/Recall Coordinator

Description of the problem and health hazard(s):

BD is conducting a voluntary medical device recall for the catalog and lot number shown above for the BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations. A small portion of this lot has been confirmed to have no additive within the tube.

As per good clinical practice, in 95% of the cases, missing additive would be detected when a visual inspection of the BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations prior to blood collection.¹ However, once blood is collected in the tubes, the clinician will be unable to determine if the tube contains additive or not. If no additive is present in the tube the sample may clot and should be rejected and recollected as per good clinical practice.

Based on publicly available scientific literature, in cases where the sample is processed without the preservative (additive) in the tube, testing has yielded reliable results if the samples were stored at room temperature for no longer than two days. If the sample was stored for more than 2 days, the result for blood alcohol determination might not be accurate (either falsely low or falsely high).²

The root cause was related to a manufacturing error and has been corrected.

Distribution of the affected lot began on March 31, 2019 and our records indicate you may have received the affected product.

Please Take the Following Actions:

1. Immediately review your inventory for the specific catalog and lot number listed above. Destroy all product subject to the recall in accordance with your institution's process for destruction.
2. Share this Urgent Medical Device Recall notification with all users of the product in your facility to ensure that they are also aware of this recall.
3. Complete the attached Customer Response/Certificate of Destruction Form and return to the BD contact noted on the form regardless of whether you have any affected material or not so that BD may acknowledge your receipt of this notification and process your product replacement, if applicable.

¹ CLSI. Collection of Diagnostic Venous Blood Specimens. 7th. Ed. CLSI standard GP41-page15, section 2.9.1 Supplies Are Gathered. Wayne, PA: Clinical and Laboratory Standards Institute; 2017. ² Wu, A. H. (2006). Tietz clinical guide to laboratory tests. St. Louis, MO: Saunders/Elsevier. Section IV- Therapeutic Drugs and Drugs of Abuse pg 1345

Report any adverse health consequences experienced with the use of this product to BD. Events may also be reported to the FDA's MedWatch Adverse Event Reporting program.

Web: MedWatch website at www.fda.gov/medwatch **Phone:** 1-800-FDA-1088 (1-800-332-1088)
Mail: MedWatch, HF-2, FDA, 5600 Fisher's Lane, Rockville, MD 20852-978

Actions Taken by BD:

1. Corrective actions have been initiated to prevent recurrence of the identified root cause.

Contact Information:

Please use the contact information provided below for complaints, adverse event reports, or questions regarding this recall.

BD Contact	US Contact Information
Customer Quality	888-237-2762 OPT 3, OPT 2 Monday – Friday 8:30am and 5:00pm (CT)

BD is committed to advancing the world of health. Our primary objectives are patient and user safety and providing you with quality products. We apologize for any inconvenience this issue may have caused you and thank you in advance for helping us to resolve this matter as quickly and effectively as possible.

Sincerely,



Aparna Jha Ahuja, MD
PG cert Hosp Management, DCH&FW, IFCCP
WW Vice President Medical Affairs, PAS



Gail Griffiths
Sr. Director, Corporate Regulatory Compliance
BD US Region

Please see July 29, 2021 BD Statement regarding Falsified June 4, 2021 Recall Notice

CUSTOMER RESPONSE/DESTRUCTION FORM

PAS-19-1461-FA

BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations

Please assist BD by promptly returning this form to: **BD Regulatory Compliance**

Email: BDRC2@bd.com or

Fax No.: Fax 312-949-0227

Facility: _____
Please use full, current facility name. Do not use initials

Street Address: _____

City: _____ **State:** _____ **Zip:** _____

Contact Person: _____

Telephone No.: _____ **Fax No.:** _____

Email Address: _____

Check all that apply:

- I have read and understood the attached notice.
- We do not have any of the affected product(s) on hand.
- I certify that I have destroyed all affected product and request replacements for the quantity shown below

Product Name	Cat. No. (Ref)	Lot No.	Units (Qty.)
BD Vacutainer® Fluoride Tubes for Blood Alcohol Determinations	367001	8215683	

Name:	
Title:	
Signature/Date:	

To further assist BD in ensuring that these notices are delivered correctly to individuals in your facility, please indicate if there is a centralized function who is responsible for managing recalls.

- No**, there is no centralized function responsible for managing recalls
- Yes**, there is a centralized function responsible for managing recalls

Function/Department Name _____

Contact Person _____

Telephone # _____ Email Address _____