



Calibration Certificate

Accredited Calibration



Cal Cert ID: 270629

Page 1 of 2

Cust. Name: Washington State Patrol
Address: 2700 116th St NE
City: Marysville
State/Zip: WA 98271 USA

Instr SN: 4100060K/651371 Out1 Min Range: 8.0000 psi
Instr Descr: CPG2400 Out1 Max Range: 17.0000 psi
Sensor SN: 4100060K/651371 Limit of Error: 0.030% R
Sensor Descr: Silicon Pressure Transducer
Procedure ID: WIO2063 Pressure Type: Absolute

Cal Date: 08/18/2025
Cal Time: 12:26 pm
Issue Date: 08/18/2025
Cal Tech: MELISSAG
Order Nbr: 61764

Test Points Values and Readings

Test Point #	Reference Value psi	Measured Values psi	Error % R	Uncertainty psi
1	8.00253	8.0026	0.0013	2.5E-4
2	9.80243	9.8024	-0.0006	3.0E-4
3	11.60227	11.6020	-0.0022	3.5E-4
4	13.40217	13.4016	-0.0042	4.1E-4
5	15.20213	15.2013	-0.0056	4.6E-4
6	17.00208	17.0010	-0.0062	5.1E-4
7	15.20205	15.2014	-0.0044	4.6E-4
8	13.40212	13.4016	-0.0039	4.1E-4
9	11.60223	11.6020	-0.0019	3.5E-4
10	9.80238	9.8023	-0.0003	3.0E-4
11	8.00249	8.0025	-0.0004	2.5E-4

Measurement Standards Used in Calibration

Std Type	Serial Number	Description	Assoc. Test Reports	Recal Date
CPD8000	10017/10016/9994	DH CPD 8000	CAL ID: 257842	01/03/2028

The combined and expanded uncertainty is reported at K=2 providing an approximate 95% confidence level.

Remarks and Notes

As Found Calibration.

As Left Calibration.

RECALIBRATION DATE: 08/18/2026

The standards and calibration program at Mensor complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 and the latest revision of the Mensor Management Manual MM 3.0.

Metrological traceability is established through accreditation by the American Association for Laboratory Accreditation (A2LA), a signatory subject to the International Laboratory Accreditation Cooperation (ILAC) Arrangement as described in Annex A of ISO 17025:2017. All calibrations are performed in ambient conditions of 23 +/- 4 deg C and 20-80% relative humidity.

The results reported herein relate only to the item calibrated. This calibration certificate shall not be reproduced, except in full, without the written approval of Mensor.

Transducer was calibrated in a horizontal position.

Prior to the as-found calibration, the zero was adjusted 0.003 psi.


Handwritten signature and date: #526, 9/16/2024, 9/16/2025



Cal Cert ID: 270629

Page 2 of 2


I, Melissa Guerrero, certify the accuracy of this Calibration Certificate:



Calibration Lab Technician

Mensor Calibration Certificates and Stickers

Mensor calibration certificates and stickers do not contain calibration intervals unless specified by the customer, as Mensor will not dictate to our customers when they should re-calibrate their instruments. The Calibration Sticker that is with this product is for your records. Due to the various location requirements of our customers, we do not affix the Calibration Sticker to the product. For Mensor products, refer to the Calibration Section of your operation manual for the recommended calibration interval.


9/16/2025 #1520